

VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N		130113	
Supervisor/Operator Performing the Verification Procedure:			
Name	Kenneth G. Cox	ID#	3527 Date 3/16/2012
A Agency		Skagway Police Department Phone # (907) 983-2232	
Instrument Location		Skagway Police Department 1st and State Street Skagway, AK 99840	
Alco S/N		4N BMB 3/22/12	
Target Value		.080 High Pressure 800 psi	
B Alco Test Values		0.079 078 BMB 3/22/12	
1st Alco		2nd Alco	
Signature			
(OVER)			

(Do Not write in the area below)


I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 5 day of April, 2012

Nikki Roth (Notary Seal Stamp)
Daniel L. Nicholas
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N

130113

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130113

MARCH 16, 2012

OPERATOR'S NAME: COX/KENNETH/G
OPERATOR'S NUMBER: 3527
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
L

O.L. #: L
DEPT/AGENCY: SKG1
CASE/REPORT: 2012-0448
TEST TYPE: U
ALCO TARGET VALUE: .000
ALCO S/N: 4N

— BREATH ANALYSIS —

.000 ADJUSTED FOR 29.47 in
ALCO TARGET .078 12:26
BLANK TEST .000 12:27
INTERNAL STANDARD VERIFIED 12:27
ALCO TU 29.47 in .079 12:27
BLANK TEST .000 12:28
SUBJECT SAMPLE .000 12:29
BLANK TEST .000 12:29
ALCO TU 29.47 in .079 12:30
BLANK TEST .000 12:31

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130113

MARCH 16, 2012
TIME 12:23

— DIAGNOSTIC CHECK —

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 02/20/01
HEATERS
SAMPLE CHAMBER: 48c
BREATH TUBE: 41c
BAROMETER: 29.47 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!"#\$%&'()*+,-./0123456789:;<=?@ABCDEFGHI
HIJKLMNOPQRSTUVWXYZ[\]^_`abcdefghijklmnopqrstuvwxyz{|}~